



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Joshua Smith  
Director

ATTESTATION


I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-003292, in compliance with the agency's Maintenance Policy and Procedures.

Signed \_\_\_\_\_

Joshua Smith

Name of Position: Director/Records Custodian



State of Oklahoma

County of Oklahoma

Signed or attested before me on 10/07/2025

by Joshua Smith.

Jordan Faith Thompson  
(Signature of notarial officer)

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing.



2/20

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

BENCH CHECK DATE: 01/04/2023

SERIAL: 80-003292

## DOCUMENTATION WITHIN SERVICE PACKET:

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- CMI Certificate of Calibration
- CMI Evaluation Form
- CMI Work Order
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis - Reference Solution (4)
- Guth Laboratories Certificate of Analysis (4)



3/20

**STATE OF OKLAHOMA**  
**BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

**CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION**


This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-003292 on 01/04/2023 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 15922080A3-009
  - BOT analysis average test result: 0.080
  - Cylinder expiration date: 08/05/2024
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR2834	DR3694	DR2455	DR3470
Concentration			
.02	.04	.10	.20
LOT #			
21340	21410	22080	22050
Manufactured Date			
8/24/2021	9/30/2021	3/7/2022	2/7/2022
Expiration Date			
8/24/2023	9/30/2023	3/7/2024	2/7/2024
Solution Commission Date			
12/13/2022	12/13/2022	12/16/2022	12/16/2022

Completed by BOT Personnel:

  
Signature/Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

  
Signature Joshua Smith  
Name

State Director  
Title



# STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

## CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT

BENCH CHECK DATE 01/04/2023	TIME PROCEDURE BEGAN 12:57:2	INSTRUMENT LOCATION 80-003292
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-003292

### TEST RESULTS

#### Diagnostics

PASS

#### Mouth Alcohol Test

PASS

#### RFI Test

PASS

#### Abort Test

PASS

#### Improper Sample Test

PASS

#### Interferent Detect Test

PASS

#### No Sample Given Test

PASS

#### Insufficient Sample Test

PASS

#### Air Blank Contamination Test

PASS

#### Range Exceeded Test

PASS


#### Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.016	0.016	0.016
.04	0.036	0.036	0.036
.10	0.098	0.097	0.098
.20	0.196	0.195	0.194

PASS

### CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 1/4/2023
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020

# Certificate of Calibration

This is to certify the calibration of **Intoxilyzer**® serial number 80-003292, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to conform to the National Highway Traffic Safety Administration (NHTSA) Standard for Devices to Measure Breath Alcohol (Federal Register, Vol.58, No.179, pp 48705-48710, Sept. 17, 1993) for accuracy and precision. Reference materials are traceable through the National Institute of Standards and Technology (NIST) to the International System of Units (SI).

Date Dec. 17, 2022

Signed

Michael Byrne

Technician

**CMI** INC.

316 East 9<sup>th</sup> Street  
Owensboro, KY 42303  
USA

Part No. 650517 Rev.A



# CMI INC.

## Service Evaluation Form

This form **MUST** be completed and enclosed with item to be serviced.  
**\*\*\* Hazardous Material Warning! – DO NOT return gas cylinder with instrument! \*\*\***

Note: Please ship items in their original shipping container.

### Contact information:

**Name** Chris Pape **Phone:** ( 405 ) 425-2460  
**Email:** christopher.pape@bot.ok.gov **Customer #** 731362 (contact Customer Service)

### Your Billing Address

**Agency Name:** Board of Tests  
**Address:** 3600 N ML King, Bldg 9  
**City:** Oklahoma City **State:** OK **Zip:** 73111  
**Credit Card/PO #:**  
**Name on Card:**  
**Expiration Date:** **CVV**

### Your Shipping Address

**Agency Name:** Board of Tests  
**Address:** 3600 N ML King, Bldg 9  
**City:** Oklahoma City **State:** OK **Zip:** 73111  
**Contact Person:** Chris Pape  
**Contact Phone:** (405) 425-2460  
**Email:** christopher.pape@bot.ok.gov

**Instrument Serial Number:** 80-003292

### Detailed Description of Problem:

Source is somewhat unstable during snap test. Flow sensor needs replaced. Display is very dim and needs replaced.

Note: For instruments not under warranty, an evaluation fee of \$83.00 (infrared and I-200) or \$43.50 (fuel cell), plus return shipping cost, will apply to all service items.

☒ I authorize all repairs up to \$500 to be performed.

After repair, add "Certificate of Calibration" for: ☐ S-D2, S-D5, and I-800 (\$19)

☐ I-200, I-240, I-300, I-400, and I-600 (\$39)

### Authorized By:

Chris Pape	Breath Testing Admin
Name (Please Print)	Title
	4/27/2022
Signature	Date

Ship item to:

**CMI, Inc.**  
**Attn: Service Dept.**  
**316 East Ninth Street**  
**Owensboro, KY 42303**



## SERVICE WORK ORDER

DATE: 5/4/2022

 316 E 9th Street / Owensboro KY 42303 / USA  
 Phone: 866-835-0690 Fax: 270-685-6268

406906

## Bill To:

 Oklahoma Board Of Tests  
 Blvd Adcu Annex Bldg #9  
 3600 Martin Luther King  
 Oklahoma City, OK 73136-3000  
 USA  
 Customer #: 731362  
 Phone: 405-425-2424  
 Fax: 405-425-2490 fax

## Ship To:

 Oklahoma Board Of Tests  
 Blvd Adcu Annex Bldg #9  
 3600 Martin Luther King  
 Oklahoma City OK 73136-3000  
 USA  
 Phone: 405-425-2424

MODEL #: 0024800K

Serial Number: 80-003292

BILL CODE: Out of Warranty

## EXTRA PARTS RCVD:

BH,KB,BATTERY COVER,GAS SHELF W/REG

## DESCRIPTION OF PROBLEM

 SOURCE IS SOMEWHAT UNSTABLE DURING SNAP TEST. FLOW SENSOR NEEDS REPLACED. DISPLAY IS VERY  
 DIM AND NEEDS REPLACED.

## WORK PERFORMED:

replaced parts listed; calibrated / final tested; 8133.16.00 software; tested with External printer/ ok;

## PARTS USED

Seq. No.	Part	Description	Quantity
100	650517	CERTIFICATE OF CALIBRATION	2.00 EA
110	441169	COVER DUST,5/8in x 1/2in	1.00 EA
120	470145	CAP,PLASTIC,.25IDx.50LG,RED	3.00 EA
130	470154	CAP,PLSTC,.406IDx.438-.562,RED	1.00 EA
131	320643	HDR,14PIN,2ROW,.100CC,4-SIDED	1.00 EA
132	140112	DISPLAY,2x20 VFD 9MM	1.00 EA
133	021297FL	PUMP ASSY,I8000	1.00 EA
135	530030	TUBING,TYGON,.375IDx.562OD	2.00 IN
140	470218	SEAL,TAMPER EVIDENT,SERVICE	1.00 EA
145	210120	SENSOR,SMD,PRESSURE	1.00 EA
150	402752	SCR,M3.5X12 POZI CSK MC,BRIGHT	1.00 EA
160	440989P03	CHASSIS,MOD	1.00 EA
170	450225	LABEL SET,8000	1.00 EA
180	400056	INSERT,EXPANSN,FOR PLASTIC,M3	9.00 EA
190	400057	INSERT,EXPANSN,FOR PLASTC,M3.5	1.00 EA
200	410278	INSERT,FOR PLASTC,M4	2.00 EA
210	410279	INSERT,EXPANSN,FOR PLASTC,M4LG	4.00 EA
215	402752	SCR,M3.5X12 POZI CSK MC,BRIGHT	4.00 EA
220	441154	MAGNET,7/16inBUTN	2.00 EA
230	450151	TAG,S/N,CMI	1.00 EA

## LABOR / TESTING

Misc Code	Description	Hours
LABR	Service Repair Labor	
LCAL	Service - Calibration Adjust	0.50
LFT	Service - Cal / Final Test	1.00

Technician Name: Dewayne Varvel

SRV4069060001





8/20

**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Dry Gas**

**ALPHA INSTRUMENT**

Intoxilyzer – Alcohol Analyzer

Model 8000


SN 80-002591

08/11/2022

14:28

Test	g/210L	Time
Air Blank	0.000	14:29
Cal Check	0.080	14:29
Air Blank	0.000	14:29
Cal Check	0.080	14:30
Air Blank	0.000	14:30
Cal Check	0.079	14:31
Air Blank	0.000	14:31
Cal Check	0.080	14:32
Air Blank	0.000	14:32
Cal Check	0.080	14:32
Air Blank	0.000	14:33
Cal Check	0.080	14:33
Air Blank	0.000	14:34
Cal Check	0.080	14:34
Air Blank	0.000	14:35
Cal Check	0.080	14:35
Air Blank	0.000	14:35
Cal Check	0.080	14:36
Air Blank	0.000	14:36
Cal Check	0.079	14:37
Air Blank	0.000	14:37
Cal Check Stats		
Average	0.0798	
Std Dev	0.0004	
Rel Std Dev (%)	0.5283	

**Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)  
0.08 g/210L = 208 PPM ÷ 2605\***

Manufacturer: <b>ILMO</b>	Manufacturer Cert. No.: <b>14879</b>	Cylinder Size: <b>105L</b>
Lot No.: <b>15922080A3</b>	Cylinder No.: <b>009</b>	Expiration Date: <b>08/05/2024</b>
<small>The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM ÷ 2605].</small>		
Signature: 		Date: <b>08/11/2022</b>
Name: <b>Christopher Pape</b>		Permit No.: <b>23020</b>





7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • [www.ilmoproducts.com](http://www.ilmoproducts.com)

## Certificate of Analysis

Certificate ID: 14879 ✓  
Part #: BAC105L080T  
Cylinder Size: 105L  
Lot Number: 15922080A3 ✓  
Expiration: 8/5/2024 ✓

ALVD  
7/29/2022  
CP

**0.080 BAC** (For the calibration of instruments used to determine breath alcohol concentration)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	208 ppm	+/-0.002 BAC(G/210L)	NDIR
Nitrogen	Balance	[5.2 ppm]	

**Distributed by:**

CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

\*Traceable to:  
NIST Research Gas Mixture - 212.6 µmol/mol  
Ethanol in Nitrogen - Serial No. SA15944

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F),

*Christopher Wright*  
Specialty Gas Lab Tech

7-19-2022  
Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory

# OFFICER'S IMPAIRED DRIVING AFFIDAVIT

10/20

Section 1

ARREST DATE 01/04/2023	ARREST TIME 1350	COUNTY # 55	CITY #	CITATION #
ARREST LOCATION BOT		CITY	COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, DELTA		DATE OF BIRTH 01/01/1985	SEX M	HEIGHT 510 WEIGHT 200
ADDRESS 3600 MLK		CITY OKLAHOMA CITY	STATE OK	ZIP CODE 73111
DRIVER LICENSE # B083463042		EXPIRATION DATE 10/31/2022	STATE OK	CLASS D RESTRICTIONS ENDORSEMENTS
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK CMV N PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST



Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

## THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000	SERIAL NUMBER 80-003292	SPECIALIST BOARD OF TESTS																																				
STANDARD Dry	TARGET VALUE 0.080	PERMIT NUMBER 23020																																				
MANUFACTURER ILMO	LOT # 15922080A3	DATE CYLINDER INSTALLED 01/04/2023 DATE CYLINDER EXPIRES 08/05/2024																																				
OBSERVATION START 1355	TEST DATE 01/04/2023	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p> SIGNATURE OF OPERATOR</p> <p>NAME PAPE, CHRISTOPHER BADGE # 0000000484 PERMIT # 0000023020 AGENCY BOARD OF TESTS</p>																																				
RESULT TABLE																																						
<table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.00</td><td>14:15</td></tr><tr><td>Subject Test</td><td>0.00</td><td>14:15</td></tr><tr><td>Breath Volume</td><td>2.097 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:16</td></tr><tr><td>Wait</td><td></td><td>14:18</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:18</td></tr><tr><td>Subject Test</td><td>0.00</td><td>14:18</td></tr><tr><td>Breath Volume</td><td>2.136 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:19</td></tr><tr><td>Cal Check</td><td>0.07</td><td>14:19</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:20</td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.00	14:15	Subject Test	0.00	14:15	Breath Volume	2.097 LITERS		Air Blank	0.00	14:16	Wait		14:18	Air Blank	0.00	14:18	Subject Test	0.00	14:18	Breath Volume	2.136 LITERS		Air Blank	0.00	14:19	Cal Check	0.07	14:19	Air Blank	0.00	14:20	 <p>Test Result BrAC g/210L <b>0.00</b></p>
Test	g/210L	Time																																				
Air Blank	0.00	14:15																																				
Subject Test	0.00	14:15																																				
Breath Volume	2.097 LITERS																																					
Air Blank	0.00	14:16																																				
Wait		14:18																																				
Air Blank	0.00	14:18																																				
Subject Test	0.00	14:18																																				
Breath Volume	2.136 LITERS																																					
Air Blank	0.00	14:19																																				
Cal Check	0.07	14:19																																				
Air Blank	0.00	14:20																																				

## What happens next?

Section 3

- The Arresting Officer is required to submit the Officer's Impaired Driving Affidavit to the State of Oklahoma.
- Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
- The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma. To confirm or change your address on file, contact Service Oklahoma.
- The Revocation and/or Disqualification will be effective 55 days from the date of the Order, including 10 days allowed for mailing. 47 O.S. §§2-116, 753, 754.
- The Order of Revocation and/or Disqualification will contain important information about your rights to appeal the revocation and/or disqualification of your driving privileges. To appeal the applicable Order of Revocation and/or Disqualification your petition must be filed with the District Court of the County in which you were arrested within 40 days of the date of the order, including 10 days allowed for mailing. 47 O.S. §§2-116, 6-211.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date:	Place (location when signed):	Signature of arresting officer:
NAME PAPE, CHRISTOPHER	AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000 PHONE 4054252460

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE



# OFFICER'S IMPAIRED DRIVING AFFIDAVIT

11/20

Section 1

ARREST DATE 01/04/2023	ARREST TIME 1350	COUNTY # 55	CITY #	CITATION #
ARREST LOCATION BOT		CITY	COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, DELTA		DATE OF BIRTH 01/01/1985	SEX M	HEIGHT 510 WEIGHT 200
ADDRESS 3600 MLK		CITY OKLAHOMA CITY	STATE OK	ZIP CODE 73111
DRIVER LICENSE # B083463042		EXPIRATION DATE 10/31/2022	STATE OK	CLASS D
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK CMV N PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST



Describe person's condition (odor, actions, etc.):

POST REPAIR MAINTENANCE, INSTALLED NEW CYLINDER

## THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000	SERIAL NUMBER 80-003292	SPECIALIST BOARD OF TESTS																																				
STANDARD Dry	TARGET VALUE 0.080	PERMIT NUMBER 23020																																				
MANUFACTURER ILMO	LOT # 15922080A3	DATE CYLINDER INSTALLED 01/04/2023 DATE CYLINDER EXPIRES 08/05/2024																																				
OBSERVATION START 1355	TEST DATE 01/04/2023	<p>A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.</p> <p>I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p> SIGNATURE OF OPERATOR</p>																																				
RESULT TABLE																																						
<table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.00</td><td>14:15</td></tr><tr><td>Subject Test</td><td>0.00</td><td>14:15</td></tr><tr><td>Breath Volume</td><td>2.097 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:16</td></tr><tr><td>Wait</td><td></td><td>14:18</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:18</td></tr><tr><td>Subject Test</td><td>0.00</td><td>14:18</td></tr><tr><td>Breath Volume</td><td>2.136 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:19</td></tr><tr><td>Cal Check</td><td>0.07</td><td>14:19</td></tr><tr><td>Air Blank</td><td>0.00</td><td>14:20</td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.00	14:15	Subject Test	0.00	14:15	Breath Volume	2.097 LITERS		Air Blank	0.00	14:16	Wait		14:18	Air Blank	0.00	14:18	Subject Test	0.00	14:18	Breath Volume	2.136 LITERS		Air Blank	0.00	14:19	Cal Check	0.07	14:19	Air Blank	0.00	14:20	  <p>Test Result BrAC g/210L <b>0.00</b></p>
Test	g/210L	Time																																				
Air Blank	0.00	14:15																																				
Subject Test	0.00	14:15																																				
Breath Volume	2.097 LITERS																																					
Air Blank	0.00	14:16																																				
Wait		14:18																																				
Air Blank	0.00	14:18																																				
Subject Test	0.00	14:18																																				
Breath Volume	2.136 LITERS																																					
Air Blank	0.00	14:19																																				
Cal Check	0.07	14:19																																				
Air Blank	0.00	14:20																																				
NAME PAPE, CHRISTOPHER		BADGE # 0000000484 PERMIT # 0000023020																																				
AGENCY BOARD OF TESTS																																						

## What happens next?

Section 3

1. The Arresting Officer is required to submit the Officer's Impaired Driving Affidavit to the State of Oklahoma.
2. Upon receipt and review of the Officer's Impaired Driving Affidavit and any associated test results, if applicable, the State of Oklahoma may revoke your Oklahoma driving privileges. If you hold a commercial driver license or were driving a commercial motor vehicle, the State of Oklahoma may disqualify your privilege to operate a commercial motor vehicle. You will continue to have driving privileges, if otherwise eligible, unless or until an Order of Revocation and/or Disqualification is issued by the State of Oklahoma.
3. The State of Oklahoma will send you the Order of Revocation and/or Disqualification to the address you have on file with Service Oklahoma. To confirm or change your address on file, contact Service Oklahoma.
4. The Revocation and/or Disqualification will be effective 55 days from the date of the Order, including 10 days allowed for mailing. 47 O.S. §§2-116, 753, 754.
5. The Order of Revocation and/or Disqualification will contain important information about your rights to appeal the revocation and/or disqualification of your driving privileges. To appeal the applicable Order of Revocation and/or Disqualification your petition must be filed with the District Court of the County in which you were arrested within 40 days of the date of the order, including 10 days allowed for mailing. 47 O.S. §§2-116, 6-211.

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct."

Section 4

Date:	Place (location when signed):	Signature of arresting officer:
NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000 PHONE 4054252460





**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000


SN 80-002591

09/23/2022

09:37

Test	g/210L	Time
Air Blank	0.000	09:37
Cal Check	0.019	09:38
Air Blank	0.000	09:39
Cal Check	0.019	09:39
Air Blank	0.000	09:40
Cal Check	0.019	09:41
Air Blank	0.000	09:41
Cal Check	0.019	09:42
Air Blank	0.000	09:42
Cal Check	0.019	09:43
Air Blank	0.000	09:44
Cal Check	0.019	09:44
Air Blank	0.000	09:45
Cal Check	0.019	09:46
Air Blank	0.000	09:46
Cal Check	0.019	09:47
Air Blank	0.000	09:47
Cal Check	0.019	09:48
Air Blank	0.000	09:49
Cal Check	0.019	09:49
Air Blank	0.000	09:50
Cal Check Stats		
Average	0.0190	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

**NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.**

Manufacturer: <b>Guth</b>	Manufactured Date: <b>8/24/2021</b>	Concentration: <b>0.02</b>
Lot No.: <b>21340</b>	Expiration Date: <b>8/24/2023</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>9/23/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

**ALPHA INSTRUMENT**

Intoxilyzer - Alcohol Analyzer


Model 8000

SN 80-002591

10/24/2022

15:45

Test	g/210L	Time
Air Blank	0.000	15:45
Cal Check	0.039	15:46
Air Blank	0.000	15:47
Cal Check	0.039	15:47
Air Blank	0.000	15:48
Cal Check	0.039	15:49
Air Blank	0.000	15:49
Cal Check	0.039	15:50
Air Blank	0.000	15:50
Cal Check	0.039	15:51
Air Blank	0.000	15:52
Cal Check	0.039	15:52
Air Blank	0.000	15:53
Cal Check	0.039	15:54
Air Blank	0.000	15:54
Cal Check	0.039	15:55
Air Blank	0.000	15:56
Cal Check	0.039	15:56
Air Blank	0.000	15:57
Cal Check	0.040	15:57
Air Blank	0.000	15:58
Cal Check Stats		
Average	0.0391	
Std Dev	0.0003	
Rel Std Dev(%)	0.8087	

<b>NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.</b>		
Manufacturer: <b>Guth</b>	Manufactured Date: <b>10/4/2021</b>	Concentration: <b>0.040</b>
Lot No.: <b>21410</b>	Expiration Date: <b>9/30/2023</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>10/24/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	





**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

**ALPHA INSTRUMENT**

Intoxilyzer - Alcohol Analyzer

Model 8000


SN 80-002591

06/27/2022

11:40

Test	g/210L	Time
Air Blank	0.000	11:40
Cal Check	0.100	11:41
Air Blank	0.000	11:41
Cal Check	0.100	11:42
Air Blank	0.000	11:43
Cal Check	0.100	11:43
Air Blank	0.000	11:44
Cal Check	0.100	11:45
Air Blank	0.000	11:45
Cal Check	0.100	11:46
Air Blank	0.000	11:47
Cal Check	0.101	11:47
Air Blank	0.000	11:48
Cal Check	0.101	11:49
Air Blank	0.000	11:49
Cal Check	0.101	11:50
Air Blank	0.000	11:50
Cal Check	0.100	11:51
Air Blank	0.000	11:52
Cal Check	0.100	11:52
Air Blank	0.000	11:53
Cal Check Stats		
Average	0.1003	
Std Dev	0.0004	
Rel Std Dev(%)	0.4815	

**NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.**

Manufacturer: <b>Guth</b>	Manufactured Date: <b>03/07/2022</b>	Concentration: <b>0.10</b>
Lot No.: <b>22080</b>	Expiration Date: <b>03/07/2024</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>06/27/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	





**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

**ALPHA INSTRUMENT**

Intoxilyzer - Alcohol Analyzer


Model 8000

SN 80-002591

06/27/2022

11:11

Test	g/210L	Time
Air Blank	0.000	11:11
Cal Check	0.199	11:12
Air Blank	0.000	11:12
Cal Check	0.200	11:13
Air Blank	0.000	11:14
Cal Check	0.200	11:14
Air Blank	0.000	11:15
Cal Check	0.200	11:16
Air Blank	0.000	11:16
Cal Check	0.201	11:17
Air Blank	0.000	11:17
Cal Check	0.200	11:18
Air Blank	0.000	11:19
Cal Check	0.201	11:19
Air Blank	0.000	11:20
Cal Check	0.200	11:21
Air Blank	0.000	11:21
Cal Check	0.200	11:22
Air Blank	0.000	11:23
Cal Check	0.201	11:23
Air Blank	0.000	11:24
Cal Check Stats		
Average	0.2002	
Std Dev	0.0006	
Rel Std Dev (%)	0.3159	

<b>NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.</b>		
Manufacturer: <b>Guth</b>	Manufactured Date: <b>02/07/2022</b>	Concentration: <b>0.20</b>
Lot No.: <b>22050</b>	Expiration Date: <b>02/07/2024</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>06/27/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21340** ✓ of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **August 25, 2021**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.0242%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **August 24, 2023** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.020 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

Testing was conducted using Cerilliant Reference Standard lot number **FN05122004** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

RCVD  
5/9/2022  
CP

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21410** ✓ of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **October 4, 2021**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.0485%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **September 30, 2023** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.040 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

**NIST Traceability:**

Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose  
values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights.  
Calibration verification is done prior to each use utilizing NIST traceable weights.

RCVD  
5/9/2022  
CP



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22080** of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **March 9, 2022**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **March 7, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

RCVD  
5/9/2022  
CP

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**RCVD  
5/9/2022  
CP

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22050** ✓ of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **February 9, 2022**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.2437%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **February 7, 2024** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.200 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

Testing was conducted using Cerilliant Reference Standard lot number **FN06231703** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Location Returned From Blanchard PD	Unit Serial # 80-003292	Date: 20/20 4/22/22
--	----------------------------	------------------------

Inventory Process

07220080A2-009

☒ Unit checked in to BOT ☒ Unlink Gas Canister from Unit ☐ Old Cylinder #:   
☒ New Unit Checked out ☐ Remove old cylinder from unit (After Lab Evaluation) ☐ Old cylinder disposed in inventory ☐ Old cylinder archived

<p><u>Check in Procedure</u></p> <p><input checked="" type="checkbox"/> Connect unit to ethernet connection 172.019.004. <u>134</u></p> <p><input checked="" type="checkbox"/> Download any stored records <u>1</u></p> <p><input checked="" type="checkbox"/> Change unit location to Serial number</p> <p><input checked="" type="checkbox"/> Update Unit Date and time</p> <p>Previous time: <u>4/27/22 1451</u></p> <p>New Time: <u>4/27/22 1453</u></p> <p><input checked="" type="checkbox"/> Execute a new Forms load on the unit</p>	<p><u>Notes from the Field</u></p> <p><input checked="" type="checkbox"/> Routine Swap <input type="checkbox"/> Issues in the Field</p>
--	---

Lab Evaluation

Breath Hose Hold Vacuum?: ☒ Test Chamber Hold Vacuum?: ☒

DVM

3µm 12708 Chamber Temp 47.00 Snap Test ok? ☒  
9µm 13120 Breath Hose Temp 45 RFI Detected? ☒

Peripheral Tests

Speaker (F5) ☒ Power LED Green (F6) ☒ Red (F7) ☒ Fan (F9) ☒  
Gas Regulator (F11) ☒ Pump (F10) ☒ Receipt paper ☒

Atmospheric Sensor

Pressure 984 Correction Factor 1.02 Flow Sensor 79 Tank Sensor 702 Is regulator close to Tank Sensor Value? ☒ Yes / ☐ No

Sent to CMI (If needed) Reason Sent Source is mildly unstable in snap test  
Date Sent 4/29/22 Flow sensor needs replaced  
Invoice Sent 7/29/22 Display is very dim  
Date Returned 12/28/22

Repairs needed:	Repairs Made:
	Ready for Bench Check? <input checked="" type="checkbox"/>

<p><u>Inventory Process</u></p> <p><input type="checkbox"/> Linked Cylinder to Unit <input type="checkbox"/> Bench Check Technician Report</p> <p><input type="checkbox"/> Changed Unit Status <input type="checkbox"/> Mock Subject Test Completed</p> <p><input type="checkbox"/> Bench Check Report Completed <input type="checkbox"/> Certificate of Calibration and Operation</p> <p><input type="checkbox"/> Service Overview Completed</p>	<p>To complete if unit sent to CMI</p> <p><input type="checkbox"/> CMI Workorder</p> <p><input type="checkbox"/> CMI Certificate of Calibration</p> <p><input type="checkbox"/> CMI Service Evaluation Form</p>
---	---