



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307

Oklahoma City, Oklahoma 73136-2307

Phone: (405) 425-2460 Fax: (405) 425-2490

www.bot.ok.gov

Dr. Jarrad Wagner, Ph.D. F-ABFT  
Chairman

Joshua Smith  
Director

**ATTESTATION**


I, the undersigned Joshua Smith (Director/Records Custodian), attest under oath pursuant to 12 O.S. §2902 this record was made, at or near the time of the occurrence of the matters set forth by or from information transmitted by a person having knowledge of those matters; was kept in the course of the regularly conducted business activity of the Board; and was made pursuant to the regularly conducted activity. The duplicate record provided is listed below:

- 1) The attached "Service Overview" is a true and accurate copy of maintenance performed on the Intoxilyzer 8000, serial number 80-004883, in compliance with the agency's Maintenance Policy and Procedures.

Signed \_\_\_\_\_

Joshua Smith

Name of Position: Director/Records Custodian



State of Oklahoma

County of Oklahoma

Signed or attested before me on 10/06/2025

by Joshua Smith.

Jordan Faith Thompson  
(Signature of notarial officer)

# Service Overview

Board of Tests for Alcohol and Drug Influence  
Enhancing public safety by ensuring the accuracy and scientific reliability  
of blood and breath testing



2/17

3600 N. Martin Luther King Bldg #9  
Oklahoma City, OK 73111  
405-425-2460

**BENCH CHECK DATE:** 11/03/2022

**SERIAL:** 80-004883

## **DOCUMENTATION WITHIN SERVICE PACKET:**

- BOT Certificate of Maintenance, Calibration and Operation
- Bench Check
- BOT Dry Gas Certificate of Analysis
- Ilmo Specialty Gases Certificate of Analysis
- Mock Subject Test
- BOT Certificate of Analysis on Certified Solutions (4)
- Guth Laboratories Certificate of Analysis (4)



3/17

**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

**CERTIFICATE of MAINTENANCE, CALIBRATION, and OPERATION**

This certifies the **Intoxilyzer 8000** referenced below, manufactured by CMI, Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, is maintained by the Board of Tests for Alcohol and Drug Influence in accordance with 40 O.A.C. 30-1-3. The instrument reported below is hereby deemed to be in compliance with Board rules via linearity accuracy examination and operational/functional procedure evaluation also referred to as a "Bench Check". The Intoxilyzer 8000 appears on the National Highway Traffic Safety Administration's Conforming Products List of Evidential Breath Alcohol Measurement Devices (82 FR 50940 *et seq*) as an approved device. In accordance with 40 O.A.C. 25-1-2, this instrument is approved and certified for use in the State of Oklahoma as an automated evidential breath-alcohol measurement device by the Board of Tests pursuant to 47 O.S. § 759.

All analyses conducted by this instrument are considered precise, valid, and in accordance with 40 O.A.C. 30-1-3(b) Operating Procedure(s). A printed breath test operator affidavit containing the seal of the Board by this instrument confirms a calibration check successfully verified the instrument's calibration at the time of test. The printed breath test affidavit also serves as proof that safeguards in place to protect the breath samples, including but not limited to Ambient Air Interference, Radio Frequency Interference, Residual Mouth Alcohol, Deficient Samples, etc. were not triggered and the test is a true and accurate scientific measurement of the person's alcohol concentration.

- Maintenance was performed on the Intoxilyzer 8000, Serial No. 80-004883 on 11/03/2022 in accordance with 40 O.A.C. 30-1-3 and internal BOT maintenance policy and procedures.
- A dry gas canister appearing on the National Highway Traffic Safety Administration's Conforming Products List of Calibrating Units for Breath Alcohol Testers (77 FR 64588 *et seq*) and approved in accordance with 40 O.A.C. 25-1-3 and BOT maintenance policy was installed.
  - Manufacturer: ILMO Unit: 208 ppm / 105 L
  - Cylinder Lot No. - Canister No.: 22722080A1-047
  - BOT analysis average test result: 0.081
  - Cylinder expiration date: 10/05/2024
- In accordance with 40 O.A.C. 25-1-2.1, the approved reference standards and equipment were used:

Simulator Model			
Guth 2100	Guth 2100	Guth 2100	Guth 2100
Serial #			
DR2834	DR3694	DR2455	DR3470
Concentration			
.02	.04	.10	.20
LOT #			
21340	21410	22080	22050
Manufactured Date			
8/24/2021	9/30/2021	3/7/2022	2/7/2022
Expiration Date			
8/24/2023	9/30/2023	3/7/2024	2/7/2024
Solution Commission Date			
11/03/2022	11/03/2022	11/03/2022	11/03/2022

Completed by BOT Personnel:

  
Signature/Name Christopher Pape

Breath-Alcohol Testing Program Administrator  
Title

  
Signature Joshua Smith  
Name

State Director  
Title



# STATE OF OKLAHOMA - BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE

## CMI INTOXILYZER BENCH CHECK PROCEDURE REPORT

BENCH CHECK DATE 11/03/2022	TIME PROCEDURE BEGAN 15:05:1	INSTRUMENT LOCATION 80-004883
INSTRUMENT TYPE Intoxilyzer	INSTRUMENT MODEL I-8000	INSTRUMENT SERIAL NUMBER 80-004883

### TEST RESULTS

#### Diagnostics

PASS

#### Mouth Alcohol Test

PASS

#### RFI Test

PASS

#### Abort Test

PASS

#### Improper Sample Test

PASS

#### Interferent Detect Test

PASS

#### No Sample Given Test

PASS

#### Insufficient Sample Test

PASS

#### Air Blank Contamination Test

PASS

#### Range Exceeded Test

PASS


#### Duplicate Breath Sampling Results

Value	Test 1	Test 2	Control Test
.02	0.018	0.017	0.019
.04	0.037	0.036	0.038
.10	0.094	0.094	0.098
.20	0.190	0.190	0.195

PASS

### CERTIFYING TECHNICIAN

I swear under penalty of perjury that in regards to the above listed instrument, I have complied with the Intoxilyzer Bench Checking Procedure approved by the State of Oklahoma.

SIGNATURE 	DATE 11/3/22
NAME PAPE, CHRISTOPHER	PERMIT # 0000023020



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Dry Gas**

ALPHA INSTRUMENT

Intoxilyzer – Alcohol Analyzer

Model 8000

SN 80-002591


10/06/2022

11:05

Test	g/210L	Time
Air Blank	0.000	11:06
Cal Check	0.081	11:06
Air Blank	0.000	11:07
Cal Check	0.081	11:07
Air Blank	0.000	11:08
Cal Check	0.082	11:08
Air Blank	0.000	11:08
Cal Check	0.081	11:09
Air Blank	0.000	11:09
Cal Check	0.081	11:10
Air Blank	0.000	11:10
Cal Check	0.081	11:10
Air Blank	0.000	11:11
Cal Check	0.081	11:11
Air Blank	0.000	11:12
Cal Check	0.081	11:12
Air Blank	0.000	11:13
Cal Check	0.082	11:13
Air Blank	0.000	11:13
Cal Check	0.081	11:14
Air Blank	0.000	11:14
Cal Check Stats		
Average	0.0812	
Std Dev	0.0004	
Rel Std Dev(%)	0.5192	

**Known Breath-Alcohol Equivalent (g/210L) Ratio of Nitrogen and Ethanol Gas (208 ppm/105 L)**

**0.08 g/210L = 208 PPM ÷ 2605\***

Manufacturer: <b>ILMO</b>	Manufacturer Cert. No.: <b>15016</b>	Cylinder Size: <b>105L</b>
Lot No.: <b>22722080A1</b>	Cylinder No.: <b>047</b>	Expiration Date: <b>10/05/2024</b>
<p>The above 105 L cylinder has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference standard meets the requirements for use as a calibrating unit in accordance with 47 O.S. 759, 40 OAC 25-1-3, and appears on the current conforming products list of calibrating units for breath alcohol testers published by the U.S. Department of Transportation in the Federal Register [77 FR 64588 et seq]. * See Dubowski, K.M. and Essary N., <i>Journal of Anal. Toxicol.</i> (1996) 20(6): pp. 484-491. [g/210L = PPM ÷ 2605].</p>		
Signature: 		Date: <b>10/06/2022</b>
Name: <b>Christopher Pape</b>		Permit No.: <b>23020</b>



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
 217-245-2183 • Fax: 217-243-7634 • [www.ilmoproducts.com](http://www.ilmoproducts.com)

## Certificate of Analysis

Certificate ID: 15016  
 Part #: BAC105L080T  
 Cylinder Size: 105L  
 Lot Number: 22722080A1  
 Expiration: 10/5/2024

RCVD  
 10/5/2022  
 CP

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:	Distributed by:
Ethanol	208 ppm	+/-0.002 BAC(G/210L)	NDIR	CMI Inc.
Nitrogen	Balance	[5.2 ppm]		316 East Ninth Street
				Owensboro, KY 42303
				Phone 866-835-0690
				<a href="http://www.alcoholtest.com">www.alcoholtest.com</a>

\*Traceable to:  
 NIST Research Gas Mixture - 212.6 µmol/mol  
 Ethanol in Nitrogen - Serial No. SA15944

Store in dry area, away from sources of heat,  
 ignition and direct sunlight. Do not allow storage  
 area to exceed 52 °C (125 °F).

*Christopher W. Wurst*  
 Specialty Gas Lab Tech

9-23-2022  
 Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

7/17

Section 1

ARREST DATE 11/03/2022	ARREST TIME 1600	COUNTY # 55	CITY #	CITATION #
ARREST LOCATION BOT		CITY		COUNTY OKLAHOMA
SUBJECT NAME BREATHTEST, DELTA		DATE OF BIRTH 01/01/1985	SEX M	HEIGHT 510 WEIGHT 200
ADDRESS 3600 MLK		CITY OKLAHOMA CITY	STATE OK	ZIP CODE 73111
DRIVER LICENSE # B083463042		EXPIRATION DATE 10/31/2022	STATE OK	CLASS D RESTRICTIONS A ENDORSEMENTS A
VEHICLE MAKE MOCK		MODEL MOCK	TAG # MOCK	STATE OK CMV N PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST


Describe person's condition (odor, actions, etc.):

ROUTINE MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000	SERIAL NUMBER 80-004883	SPECIALIST BOARD OF TEST																																				
STANDARD Dry	TARGET VALUE 0.080	PERMIT NUMBER 23020																																				
MANUFACTURER ILMO	LOT # 22722080A1	DATE CYLINDER INSTALLED 11/03/2022 DATE CYLINDER EXPIRES 10/05/2024																																				
OBSERVATION START 1615	TEST DATE 11/03/2022	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.																																				
RESULT TABLE																																						
<table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.00</td><td>16:50</td></tr><tr><td>Subject Test</td><td>0.00</td><td>16:51</td></tr><tr><td>Breath Volume</td><td>2.117 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>16:51</td></tr><tr><td>Wait</td><td></td><td>16:53</td></tr><tr><td>Air Blank</td><td>0.00</td><td>16:53</td></tr><tr><td>Subject Test</td><td>0.00</td><td>16:54</td></tr><tr><td>Breath Volume</td><td>2.054 LITERS</td><td></td></tr><tr><td>Air Blank</td><td>0.00</td><td>16:54</td></tr><tr><td>Cal Check</td><td>0.08</td><td>16:54</td></tr><tr><td>Air Blank</td><td>0.00</td><td>16:55</td></tr></tbody></table>		Test	g/210L	Time	Air Blank	0.00	16:50	Subject Test	0.00	16:51	Breath Volume	2.117 LITERS		Air Blank	0.00	16:51	Wait		16:53	Air Blank	0.00	16:53	Subject Test	0.00	16:54	Breath Volume	2.054 LITERS		Air Blank	0.00	16:54	Cal Check	0.08	16:54	Air Blank	0.00	16:55	I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.  SIGNATURE OF OPERATOR
Test	g/210L	Time																																				
Air Blank	0.00	16:50																																				
Subject Test	0.00	16:51																																				
Breath Volume	2.117 LITERS																																					
Air Blank	0.00	16:51																																				
Wait		16:53																																				
Air Blank	0.00	16:53																																				
Subject Test	0.00	16:54																																				
Breath Volume	2.054 LITERS																																					
Air Blank	0.00	16:54																																				
Cal Check	0.08	16:54																																				
Air Blank	0.00	16:55																																				
NAME PAPE, CHRISTOPHER																																						
BADGE # 0000000484		PERMIT # 0000023020																																				
AGENCY BOARD OF TESTS																																						



Test Result  
BrAC g/210L  
0.00

EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED. Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below.

You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct." Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER	AGENCY BOARD OF TESTS		
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

OTHER WITNESSES:

Section 5

1	NAME	TITLE	ADDRESS	PHONE
2	NAME	TITLE	ADDRESS	PHONE

OFFICER/DPS COPY

BOT Form AFF01 09/2021



# OFFICER'S AFFIDAVIT AND NOTICE OF REVOCATION/DISQUALIFICATION

8/17

Section 1

ARREST DATE 11/03/2022		ARREST TIME 1600		COUNTY # 55	CITY #	CITATION #	
ARREST LOCATION BOT				CITY		COUNTY OKLAHOMA	
SUBJECT NAME BREATHTEST, DELTA				DATE OF BIRTH 01/01/1985		SEX M	HEIGHT 510
						WEIGHT 200	
ADDRESS 3600 MLK				CITY OKLAHOMA CITY		STATE OK	ZIP CODE 73111
DRIVER LICENSE # B083463042				EXPIRATION DATE 10/31/2022		STATE OK	CLASS D
						RESTRICTIONS	ENDORSEMENTS A
VEHICLE MAKE MOCK				MODEL MOCK		TAG # MOCK	STATE OK
						CMV N	PLACARD N

On the above date, time, and location, the above named person was arrested, and I had reasonable grounds to believe the person was driving or in actual physical control of a motor vehicle upon a public road, highway, street, turnpike, public parking lot, or other public place, or upon a private road, street, alley or lane which provides access to one or more single or multi-family dwelling within this state while under the influence of alcohol and/or other intoxicating substances as prohibited by law.

(Describe driving behavior or circumstances):

MOCK SUBJECT TEST

Describe person's condition (odor, actions, etc.):

ROUTINE MAINTENANCE, INSTALLED NEW CYLINDER

THE PERSON WAS INFORMED OF THE IMPLIED CONSENT TEST REQUEST

BREATH-ALCOHOL ANALYSIS RECORD, REPORT OF THE PERSON NAMED IN SECTION 1 ABOVE, and LOG OF TEST AND MAINTENANCE RECORD.

Section 2

EQUIPMENT I-8000		SERIAL NUMBER 80-004883	SPECIALIST BOARD OF TEST																																					
STANDARD Dry		TARGET VALUE 0.080	PERMIT NUMBER 23020																																					
MANUFACTURER ILMO		LOT # 22722080A1	DATE CYLINDER INSTALLED 11/03/2022	DATE CYLINDER EXPIRES 10/05/2024																																				
OBSERVATION START 1615	TEST DATE 11/03/2022	A 15-minute (or greater) deprivation/observation period was observed prior to and continuing to the time of test administration.																																						
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.00</td><td>16:50</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>16:51</td></tr> <tr><td>Breath Volume</td><td>2.117 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>16:51</td></tr> <tr><td>Wait</td><td></td><td>16:53</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>16:53</td></tr> <tr><td>Subject Test</td><td>0.00</td><td>16:54</td></tr> <tr><td>Breath Volume</td><td>2.054 LITERS</td><td></td></tr> <tr><td>Air Blank</td><td>0.00</td><td>16:54</td></tr> <tr><td>Cal Check</td><td>0.08</td><td>16:54</td></tr> <tr><td>Air Blank</td><td>0.00</td><td>16:55</td></tr> </tbody> </table>		Test	g/210L	Time	Air Blank	0.00	16:50	Subject Test	0.00	16:51	Breath Volume	2.117 LITERS		Air Blank	0.00	16:51	Wait		16:53	Air Blank	0.00	16:53	Subject Test	0.00	16:54	Breath Volume	2.054 LITERS		Air Blank	0.00	16:54	Cal Check	0.08	16:54	Air Blank	0.00	16:55	<p style="text-align: center;">I administered said Breath-Alcohol Test in compliance with the applicable operating procedure of the Board of Tests.</p> <p style="text-align: center;"></p> <p style="text-align: center;">SIGNATURE OF OPERATOR</p>		
		Test	g/210L	Time																																				
		Air Blank	0.00	16:50																																				
		Subject Test	0.00	16:51																																				
		Breath Volume	2.117 LITERS																																					
		Air Blank	0.00	16:51																																				
		Wait		16:53																																				
		Air Blank	0.00	16:53																																				
		Subject Test	0.00	16:54																																				
		Breath Volume	2.054 LITERS																																					
Air Blank	0.00	16:54																																						
Cal Check	0.08	16:54																																						
Air Blank	0.00	16:55																																						
<p>NAME PAPE, CHRISTOPHER</p>																																								
BADGE # 0000000484			PERMIT # 0000023020																																					
AGENCY BOARD OF TESTS																																								



Test Result  
BrAC g/210L  
**0.00**

**EFFECTIVE THIRTY (30) DAYS FROM DATE SERVED.** Notice of Revocation is given that your driver's license is revoked or denied for a period of one hundred eighty (180) days or more, if you refused or failed the state's test(s). In addition, your commercial driving privilege in this state will be disqualified if you refused or failed the state's test(s) (operating a CMV or CDL holder) OR a test result of .04 or more while operating a commercial motor vehicle. You may appeal any Departmental action against your driver license as a result of this arrest by filing a petition in the District Court of the County in which you were arrested. Your petition must be filed within thirty (30) days of the date listed below. You may be eligible to participate in the Impaired Driver Accountability Program (IDAP) administered by the Department of Public Safety (DPS). Completion of IDAP may prevent a revocation as a result of this arrest from appearing on your driving record. Participation in IDAP may reduce the amount of fees you will be required to pay to the State. You must request IDAP within thirty (30) days of the date listed below. Mail the request for IDAP to the Department of Public Safety at P.O. Box 11415, Oklahoma City, OK 73136.

Section 3

**SERVING OF NOTICE:** I personally hand delivered a copy of the above containing the Notice of Revocation and the Notice of Department Impaired Driving Accountability Program to the person named in Section 1 above.

Date served: \_\_\_\_\_ Signature of officer: \_\_\_\_\_ Badge #: \_\_\_\_\_

In accordance with Title 12 O.S. Section 426, "I state under penalty of perjury under the laws of Oklahoma that the foregoing is true and correct." Section 4

Date: \_\_\_\_\_ Place (location when signed): \_\_\_\_\_ Signature of arresting officer: \_\_\_\_\_

NAME PAPE, CHRISTOPHER		AGENCY BOARD OF TESTS	
BADGE 0000000484	AGENCY ADDRESS INTER-AGENCY	ZIP CODE 00000	PHONE 4054252460

ARRESTEE'S COPY

BOT Form AFF02 09/2021





**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000


SN 80-002591

09/23/2022

09:37

Test	g/210L	Time
Air Blank	0.000	09:37
Cal Check	0.019	09:38
Air Blank	0.000	09:39
Cal Check	0.019	09:39
Air Blank	0.000	09:40
Cal Check	0.019	09:41
Air Blank	0.000	09:41
Cal Check	0.019	09:42
Air Blank	0.000	09:42
Cal Check	0.019	09:43
Air Blank	0.000	09:44
Cal Check	0.019	09:44
Air Blank	0.000	09:45
Cal Check	0.019	09:46
Air Blank	0.000	09:46
Cal Check	0.019	09:47
Air Blank	0.000	09:47
Cal Check	0.019	09:48
Air Blank	0.000	09:49
Cal Check	0.019	09:49
Air Blank	0.000	09:50
Cal Check Stats		
Average	0.0190	
Std Dev	0.0000	
Rel Std Dev (%)	0.0000	

**NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.**

Manufacturer: <b>Guth</b>	Manufactured Date: <b>8/24/2021</b>	Concentration: <b>0.02</b>
Lot No.: <b>21340</b>	Expiration Date: <b>8/24/2023</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>9/23/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

ALPHA INSTRUMENT

Intoxilyzer - Alcohol Analyzer

Model 8000


SN 80-002591

10/24/2022

15:45

Test	g/210L	Time
Air Blank	0.000	15:45
Cal Check	0.039	15:46
Air Blank	0.000	15:47
Cal Check	0.039	15:47
Air Blank	0.000	15:48
Cal Check	0.039	15:49
Air Blank	0.000	15:49
Cal Check	0.039	15:50
Air Blank	0.000	15:50
Cal Check	0.039	15:51
Air Blank	0.000	15:52
Cal Check	0.039	15:52
Air Blank	0.000	15:53
Cal Check	0.039	15:54
Air Blank	0.000	15:54
Cal Check	0.039	15:55
Air Blank	0.000	15:56
Cal Check	0.039	15:56
Air Blank	0.000	15:57
Cal Check	0.040	15:57
Air Blank	0.000	15:58
Cal Check Stats		
Average	0.0391	
Std Dev	0.0003	
Rel Std Dev(%)	0.8087	

**NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.**

Manufacturer: <b>Guth</b>	Manufactured Date: <b>10/4/2021</b>	Concentration: <b>0.040</b>
Lot No.: <b>21410</b>	Expiration Date: <b>9/30/2023</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>10/24/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	





**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

**ALPHA INSTRUMENT**

Intoxilyzer - Alcohol Analyzer


Model 8000

SN 80-002591

06/27/2022

11:40

Test	g/210L	Time
Air Blank	0.000	11:40
Cal Check	0.100	11:41
Air Blank	0.000	11:41
Cal Check	0.100	11:42
Air Blank	0.000	11:43
Cal Check	0.100	11:43
Air Blank	0.000	11:44
Cal Check	0.100	11:45
Air Blank	0.000	11:45
Cal Check	0.100	11:46
Air Blank	0.000	11:47
Cal Check	0.101	11:47
Air Blank	0.000	11:48
Cal Check	0.101	11:49
Air Blank	0.000	11:49
Cal Check	0.101	11:50
Air Blank	0.000	11:50
Cal Check	0.100	11:51
Air Blank	0.000	11:52
Cal Check	0.100	11:52
Air Blank	0.000	11:53
Cal Check Stats		
Average	0.1003	
Std Dev	0.0004	
Rel Std Dev (%)	0.4815	

<b>NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.</b>		
Manufacturer: <b>Guth</b>	Manufactured Date: <b>03/07/2022</b>	Concentration: <b>0.10</b>
Lot No.: <b>22080</b>	Expiration Date: <b>03/07/2024</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 		Date: <b>06/27/2022</b>
Name <b>Christopher Pape</b>		Permit No.: <b>23020</b>



**STATE OF OKLAHOMA  
BOARD OF TESTS FOR ALCOHOL AND DRUG INFLUENCE**

Dr. Jarrad Wagner, Ph.D., F-ABFT  
Chairman

Post Office Box 36307  
Oklahoma City, Oklahoma 73136-2307  
Phone: (405) 425-2460 Fax: (405) 425-2490  
www.bot.ok.gov

Joshua Smith  
Director

**Certificate of Analysis – Reference Solution**

**ALPHA INSTRUMENT**

Intoxilyzer - Alcohol Analyzer


Model 8000

SN 80-002591

06/27/2022

11:11

Test	g/210L	Time
Air Blank	0.000	11:11
Cal Check	0.199	11:12
Air Blank	0.000	11:12
Cal Check	0.200	11:13
Air Blank	0.000	11:14
Cal Check	0.200	11:14
Air Blank	0.000	11:15
Cal Check	0.200	11:16
Air Blank	0.000	11:16
Cal Check	0.201	11:17
Air Blank	0.000	11:17
Cal Check	0.200	11:18
Air Blank	0.000	11:19
Cal Check	0.201	11:19
Air Blank	0.000	11:20
Cal Check	0.200	11:21
Air Blank	0.000	11:21
Cal Check	0.200	11:22
Air Blank	0.000	11:23
Cal Check	0.201	11:23
Air Blank	0.000	11:24
Cal Check Stats		
Average	0.2002	
Std Dev	0.0006	
Rel Std Dev (%)	0.3159	

<b>NIST Traceable Breath-Alcohol Equivalent (g/210L +/- 3%) Ratio of ethanol and water free of test interfering substances.</b>		
Manufacturer: <b>Guth</b>	Manufactured Date: <b>02/07/2022</b>	Concentration: <b>0.20</b>
Lot No.: <b>22050</b>	Expiration Date: <b>02/07/2024</b>	Equipment Model: <b>Guth 2100</b>
The above solution has been analyzed via Infrared Spectroscopy utilizing the Intoxilyzer 8000 and was performed in accordance with the procedures prescribed by the State Director of Tests pursuant to 40 OAC 30-1-3. The above NIST traceable reference solution meets the requirements for use pursuant to 40 OAC 25-1-2.1. Approved breath-alcohol measurement equipment and reference solutions. Equipment used appears on the current or supplemented Conforming Products List of Calibrating Units for Breath Alcohol Testers (72 FR 34747), published by the National Highway Traffic Safety Administration.		
Signature 	Date: <b>06/27/2022</b>	
Name <b>Christopher Pape</b>	Permit No.: <b>23020</b>	



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21340** ✓ of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **August 25, 2021**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.0242%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **August 24, 2023** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.020 g/210L** +/- 3%. ✓

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number **FN05122004** whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

RCVD  
5/9/2022  
CP

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21410** ✓ of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **October 4, 2021**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.0485%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **September 30, 2023** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.040 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number **FN08211802** whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

RCVD  
5/9/2022  
CP



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22080** of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **March 9, 2022**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **March 7, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number **FN03052002** whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*

RCVD  
5/9/2022  
CP

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

**CERTIFICATE OF ANALYSIS**

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22050** ✓ of  
Alcohol Reference Solution for Simulator were analyzed by  
gas chromatography on **February 9, 2022**, using a Perkin Elmer Gas  
Chromatograph Autosystem XL S/N: 610N9030209, and found to contain  
**0.2437%** (w/vol) ethyl alcohol. The expiration date for this lot  
number is **February 7, 2024** at 11:59 PM. ✓

When used in a calibrated Simulator, operating at  
34°C +/- .2°C, this solution will give a breath alcohol  
analysis instrument reading of **0.200 g/210L +/- 3%**. ✓

The alcohol and water used in this solution were  
free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

Testing was conducted using Cerilliant Reference Standard lot number **FN06231703** whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

RCVD  
5/9/2022  
CP



<u>Location Returned From</u> Jenks PD	<u>Unit Serial #</u> 80-004883	<div style="text-align: right; color: red;">17/17</div> <u>Date:</u> 11/2/22
---	-----------------------------------	---

Inventory Process

Old Cylinder #: 22620080A2-001

☒ Unit checked in to BOT  
☒ New Unit Checked out

☒ Unlink Gas Canister from Unit  
☒ Remove old cylinder from unit (After Lab Evaluation)

☒ Old cylinder disposed in inventory  
☒ Old cylinder archived

<u>Check in Procedure</u> <input checked="" type="checkbox"/> Connect unit to ethernet connection 172.019.004. 137 <input checked="" type="checkbox"/> Download any stored records 17 <input checked="" type="checkbox"/> Change unit location to Serial number <input checked="" type="checkbox"/> Update Unit Date and time Previous time: 11/3/22 1027 New Time: 11/3/22 1030 <input checked="" type="checkbox"/> Execute a new Forms load on the unit	<u>Notes from the Field</u> <input checked="" type="checkbox"/> Routine Swap <input type="checkbox"/> Issues in the Field
--	--

Lab Evaluation  
 Breath Hose Hold Vacuum? ☒      Test Chamber Hold Vacuum? ☒

DVM  

3µm 12884	Chamber Temp 47.00	Snap Test ok? <input checked="" type="checkbox"/>
9µm 13375	Breath Hose Temp 45	RFI Detected? <input checked="" type="checkbox"/>

Peripheral Tests  

Speaker (F5) <input checked="" type="checkbox"/>	Power LED Green (F6) <input checked="" type="checkbox"/>	Red (F7) <input checked="" type="checkbox"/>	Fan (F9) <input checked="" type="checkbox"/>
Gas Regulator (F11) <input checked="" type="checkbox"/>	Pump (F10) <input checked="" type="checkbox"/>	Receipt paper <input checked="" type="checkbox"/>	

<u>Atmospheric Sensor</u> Pressure 965      Correction Factor 1.04	<u>Flow Sensor</u> Resistance 180	<u>Tank Sensor</u> Value 970
---	--------------------------------------	---------------------------------

Is regulator close to Tank Sensor Value?  
☒ Yes / ☐ No

Sent to CMI (If needed) Date Sent _____ Invoice Sent _____ Date Returned _____	Reason Sent
---	-------------

<u>Repairs needed:</u>	<u>Repairs Made:</u>          <div style="text-align: right; margin-top: 20px;">           Ready for Bench Check? <input checked="" type="checkbox"/> </div>
------------------------	--

<u>Inventory Process</u> <input type="checkbox"/> Linked Cylinder to Unit <input type="checkbox"/> Changed Unit Status <input type="checkbox"/> Bench Check Report Completed	<div style="display: flex; justify-content: space-between;"> <div> <input type="checkbox"/> Bench Check Technician Report  <input type="checkbox"/> Mock Subject Test Completed  <input type="checkbox"/> Certificate of Calibration and Operation  <input type="checkbox"/> Service Overview Completed           </div> <div>             To complete if unit sent to CMI  <input type="checkbox"/> CMI Workorder  <input type="checkbox"/> CMI Certificate of Calibration _____  <input type="checkbox"/> CMI Service Evaluation Form           </div> </div>
---	---

For BOT Internal Use Only

Intoxilyzer Check-in sheet Version 2.0